

Department of Forensic Science  
Breath Alcohol Section  
Instrument Location History  
1/1/1990 To 4/23/2015

*[Handwritten signature]*  
*mmB*

Instrument Serial Number: 010548

As of 23-Apr-15

Start Date	End Date	Agency Name
22-Apr-15		DFS Central Lab
18-Nov-14	22-Apr-15	Accomack County SO
29-Oct-14	18-Nov-14	DFS Central Lab
30-May-14	29-Oct-14	Bedford County SO
22-May-14	30-May-14	DFS Central Lab
12-Jan-09	22-May-14	King & Queen Sheriff's Office
23-Jun-08	12-Jan-09	DFS Central Lab

ANS  
mmB

# INTOX EC/IR II

## Quality Assurance Worksheet

Instrument Serial Number 010548 Worksheet Start Date 4/22/2015

Location Accomack County

Address 23371 Front St., Accomac, VA 23301

DFS Technician Heather Stanton License No. 18910

☐ Laboratory ☒ On-Site

Site Specification: No detrimental environmental conditions exist. ☒

Instrument Barometer (mm HG) 752 Reference Barometer (mm HG) 756

Reference Barometer(RB)Serial # 009111 RB Calibration Due 5/19/2015

### Measurement Assurance Check

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.302
0.300	0.297	0.288	0.305	Sample 2	0.301
Precision		sample min	sample max	Sample 3	0.301
0.001		0.301	0.302		

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.100
0.100	0.099	0.096	0.102	Sample 2	0.100
Precision		sample min	sample max	Sample 3	0.100
0		0.100	0.100		


Dry gas standard Lot No. (with tank no.) AG425201-24

☐ Replaced dry gas standard (+O-ring)

☐ Installed at Location

☒ Removed to DFS-Central

Supplies  
Mouthpieces  
Certificates of Analysis  
Operator Worksheet  
Other:



Notes:

4/22/2015

Instrument Serial Number

010548

Certification Date

04/23/15

☐ Calibrated☐ Certified☒ Measurement Assurance Check☒ Instrument Test☐ Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.250		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.150		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.080		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.020		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

DFS Technician

Heath W. Stach

Date

4/23/15

Issuing Analyst

Michelle Bullis

Date

4/24/15




COMMONWEALTH OF VIRGINIA  
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS  
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

AMS  
mmB

NAME OF ACCUSED <b>INSTRUMENT, TEST,</b>	NAME OF COURT <b>DFS</b>
---	-----------------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY <b>STANTON, HEATHER, M</b>		AGENCY <b>DFS Central Lab</b>
DFS LICENSE NUMBER <b>18910</b>	LICENSE EXPIRES <b>10/01/2016</b>	DATE TEST CONDUCTED <b>04/22/2015</b>
TEST EQUIPMENT NUMBER <b>010548</b>		

RESULTS: TIME SAMPLE TAKEN 10:07 EDT

SAMPLE'S ALCOHOL CONTENT 0.00 GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS \_\_\_\_\_ DAY OF \_\_\_\_\_, 20\_\_\_\_\_.

BREATH TEST OPERATOR
----------------------

☐ I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS \_\_\_\_\_ SUBJECT'S SIGNATURE

☐ SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS \_\_\_\_\_ OPERATOR'S SIGNATURE



# IntoxNet MIS Report

Report Generated 22 Apr 2015 at 10:30

hms  
mmB

## Test Results

Instrument Serial Number 010548

Test # 000429 Subject Test

Test Location 1 Department of  
Test Date 22 Apr 2015

Test Location 2 Forensic Science  
Test Time 10:00  
Remote/Local Local

Test Location 3  
System Check Passed

Operator's Last Name STANTON  
Agency DFS Central Lab

Operator's First Name HEATHER

Operator's Middle Initial M

Card Serial Number 118910

Effective Date 10/01/2014

License Number 18910

Expiration Date 10/01/2016

Subject's Last Name INSTRUMENT

Subject's First Name TEST

Subject's Middle Initial

Subject's Date of Birth 00/00/0000

Subject's Sex Male

Driver's License Number

Driver's License Expiration 00/00/0000

Driver's License State

Court Name DFS

End Date 22 Apr 2015 End Time 10:08

Result Time 10:07

Result Date 22 Apr 2015 Result 0.00

Data Type DIAG

Sample Value Pass

Sample Time 10:01

Data Type BLK

Sample Value 0.000

Sample Time 10:01

Data Type CHK

Sample Value 0.101

Sample Time 10:02

Data Type BLK

Sample Value 0.000

Sample Time 10:04

Data Type SUBJ

Sample Value 0.000

Sample Time 10:04

Data Type BLK

Sample Value 0.000

Sample Time 10:05

Data Type SUBJ

Sample Value 0.000

Sample Time 10:07

Data Type BLK

Sample Value 0.000

Sample Time 10:08

Standard Type Dry Gas Std

Standard Value 0.099

Standard Lot Number AG425201-24

Standard Expiration Date 09/09/2016

Tank Pressure 1061

Barometric Pressure 752 mmHg

Blow Sample Number 1 Blow Duration 3.16 sec

Blow Volume 1738 cc End-of-Blow Time 10:04

Blow Sample Number 2 Blow Duration 3.51 sec

Blow Volume 1740 cc End-of-Blow Time 10:07

Tamper Evident Stamp 6cf35e3d

Test Status Code 0

Test Status Success

AMS

mmB

**Intox EC/IR-II: Accuracy Check**

*Department of Forensic Science*

Serial Number: 010548      Test Number: 430

Test Date: 04/22/2015      Test Time: 10:09 EDT

Dry Gas Target: 0.099

Lot Number: AG425201-24      Exp Date: 09/09/2016

Tank Pressure: 1061 psi      Barometric Pressure: 752 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	10:10
CHK	0.100	10:10
BLK	0.000	10:12
CHK	0.100	10:13
BLK	0.000	10:14
CHK	0.100	10:15

Test Status: *Success*

Calibration CRC: D11EF1A3

AMS

mmB

**Intox EC/IR-II: Accuracy Check**

*Department of Forensic Science*

Serial Number: 010548      Test Number: 431

Test Date: 04/22/2015      Test Time: 10:17 EDT

Dry Gas Target: 0.297

Lot Number: AG506501-09      Exp Date: 03/06/2017

Tank Pressure: 183 psi      Barometric Pressure: 752 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	10:18
CHK	0.302	10:19
BLK	0.000	10:21
CHK	0.301	10:21
BLK	0.000	10:23
CHK	0.301	10:24

Test Status: *Success*

Calibration CRC: D11EF1A3